14.06.2018 14:35

## Model: UTS ergo line XLT Reference no.: 29-061467-001

Correct use: passive storage



Product group:	safety storage cabinet
Fire resistance (FR):	Type 90
Catalogue page:	120
Doors can be opened with minimum effort, design:	wing door
Door technology:	classical
In case of fire, automatic and blockage-free self- closing via thermocouple:	yes
Access control with cylinder lock:	✓
Enhanced ventilation concept - ventilation in every cabinet level:	✓
Visual inspection of ventilation cut-off flaps:	✓
Outer width [mm]:	1400 - (55.12 inch)
Outer depth [mm]:	593 - (23.35 inch)
Outer height [mm]:	631 - (24.84 inch)
Base height [mm]:	35
Base version:	classical, with slots and floor extraction



14.06.2018 14:35

Material outer carcass:	sheet steel, powder-coated
Gross weight [kg]:	220 - (485.02 lbs)
Center partition wall:	✓
Number of sides:	2
Inner width [mm]:	790 - (31.10 inch)
Inner depth [mm]:	463 - (18.23 inch)
Inner height [mm]:	480 - (18.90 inch)
Material inner carcass:	decor panels
Maximum volume of largest individual container - left side [l]:	19 - (5.02 gal)
Maximum volume of largest individual container - right side [l]:	12 - (3.17 gal)
Collection volume - left side [l]:	21.3 - (5.63 gal)
Collection volume - right side [l]:	13.2 - (3.49 gal)
Bottom tray interior:	bottom tray
Number of bottom trays:	1
Perforated sheet interior:	perforated sheet insert
Number of perforated sheet inserts:	1
Earthing connection according to TRGS 727 available:	2
Number of earthing connections:	2
Exhaust air connection included:	NW 75
Standing surface material:	sheet steel, powder-coated



14.06.2018 14:35

Load-bearing capacity per standing surface with uniformly distributed load - left side [kg]:	60 - (132.28 lbs)
Load-bearing capacity per standing surface with uniformly distributed load - right side [kg]:	60 - (132.28 lbs)
Adjustment height 1 left [mm]:	386 - (15.20 inch)
Innenbreite [mm] 2:	490 - (19.29 inch)
Mark of approval by TÜV:	GS mark
Type-tested by recognised test institutes according to:	DIN EN 14470-1
TÜV-tested according to:	DIN EN 14470-1, EK5/AK4 09-10
Further TÜV testing (1):	DIN EN 14727
Further TÜV testing (2):	TRGS 510 Annex 3
Added value mark of approval by TÜV:	High Quality
Declaration of conformity:	CE