



Product Service

# CERTIFICATE

No. Z1A 012906 0514 Rev. 00

**Holder of Certificate:** **DÜPERTHAL SICHERHEITSTECHNIK  
GMBH & Co. KG**



**DÜPERTHAL**

Frankenstraße 3  
63791 Karlstein  
GERMANY

**Certification Mark:**



**Product:** **Safety cabinets**

**Tested according to:** PPP 52125B:2019  
DIN EN 16121:2017  
DIN EN 16122:2012  
AfPS GS 2019:01 PAK

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition, the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 713165795-100

**Valid until:** 2026-06-01

**Date,** 2021-07-23

*Horst Kristen*

( Horst Kristen )

# CERTIFICATE

No. Z1A 012906 0514 Rev. 00

**Model(s):** ACID C pro  
ACID U pro  
ACID B pro

## Kenndaten:

<b>ACID C pro</b>	
Model size	approx. size with socket in mm (width x depth x height)
S	600 x 609 x 1385
XS	1200 x 609 x 1385
M	600 x 609 x 2045
XL	1200 x 609 x 2045

<b>ACID U pro</b>	
Model size	approx. size without socket in mm (width x depth x height)
S	600 x 609 x 596
L	1100 x 609 x 596

<b>ACID B pro</b>	
Model size	approx. size without socket in mm (width x depth x height)
S	600 x 609 x 770
L	1100 x 609 x 770

Class of fire resistance: FWF 0

A detailed description of the specifications can be found in the test report.